

Certified Reference Material - Turbidity, potable water (PotableWatR TM)

Art. ID ERA-699
Unit each
Deliverydetails No Dangerous Good /not restricted

Description

One 24 mL screw-cap vial yields up to 1 liter after dilution. Use with nephelometric methods. The certified values vary from batch to batch. Please ask for the certificate.

Text/Information	Analyte/Parameter	CAS number	Concentration/Value	Unit	Method	Source
	Turbidity		0,5 - 8,0	NTU		