

**Certified Reference Material - Organochlorine Pesticides, priority pollutant (PriorityPollutnT™)**

Art. ID ERA-713  
Unit each  
Deliverydetails No Dangerous Good /not restricted

**Description**

One 2 mL flame-sealed ampule yields up to 2 liters after dilution. Contains contains a subset of the analytes listed below at 0.5-20 µg/L. The certified values vary from batch to batch. Please ask for the certificate.

Text/Information	Analyte/Parameter	CAS number	Concentration/Value	Unit	Method	Source
	4,4'-DDD	[72-54-8]	2,0 -10,0	µg/L		
	4,4'-DDE	[72-55-9]	2,0 -10,0	µg/L		
	4,4'-DDT	[50-29-3]	1,0 - 10	µg/L		
	Aldrin	[309-00-2]	0,5 - 15	µg/L		
	Dieldrin	[60-57-1]	1,0 -13,0	µg/L		
	Endosulfan-I (alpha)	[959-98-8]	1,0 - 4,0	µg/L		
	Endosulfan-II (beta)	[33213-65-9]	4,0 - 20,0	µg/L		
	Endosulfan-sulfate	[1031-07-8]	2,0 - 20,0	µg/L		
	Endrin	[72-20-8]	2,0 - 20,0	µg/L		
	Endrin aldehyde	[7421-93-4]	4,0 - 20,0	µg/L		
	Endrin ketone	[53494-70-5]	2,0 -10,0	µg/L		
	Heptachlor	[76-44-8]	1,0 - 10	µg/L		
	Heptachlor Epoxide (bet a)	[1024-57-3]	1,0 - 10	µg/L		
	Methoxychlor	[72-43-5]	2,0 - 15,0	µg/L		
	alpha-HCH	[319-84-6]	2,0 - 15,0	µg/L		
	cis-Chlordane (alpha)	[5103-71-9]	1 - 9,8	µg/L		
	beta-HCH	[319-85-7]	2,0 - 15,0	µg/L		
	delta-HCH	[319-86-8]	2,0 - 15,0	µg/L		
	gamma-HCH (Lindane)	[58-89-9]	2,0 - 15,0	µg/L		
	trans-Chlordane (gamma)	[5103-74-2]	1,2 - 7,8	µg/L		