

Certified Reference Material - Low-Level Total Residual Chlorine

Art. ID	ERA-917
Unit	each
Deliverydetails	No Dangerous Good /not restricted

Description

Designed for testing at low µg/L levels. One 2 mL flame-sealed ampule yields up to 2 liters after dilution. Use with sensitive titrimetric or colorimetric methods. The certified values vary from batch to batch. Please ask for the certificate.

Text/Information	Analyte/Parameter	CAS number	Concentration/Value	Unit	Method	Source
	Total residual chlorine		20 - 250	µg/L		