

## **Proficiency Testing - Demand, waste water**

Art. ID	ERA-578
Unit	each
Deliverydetails	No Dangerous Good /not restricted

### Description

One 15 mL screw-cap vial yields up to 2 liters after dilution /// Methods: EPA 405.1, EPA 410.1, EPA 410.2, EPA 410.3, EPA 410.4, EPA 415.1, EPA 415.2, EPA 9060, SM 18/19thED 5310B, SM 18/19thED 5310C, SM 18/19thED 5310D, SM 18th ED 5210 B, SM 18th ED 5520 B

Text/Information	Analyte/Parameter	CAS number	Concentration/Value	Unit	Method	Source
	5-day BOD Biochemical Oxygen Demand (BOD)		15 - 250	mg/L		
	Carbonous BOD (CBOD)		15 - 250	mg/L		
	Chemical Oxygen Demand (COD)		30 - 250	mg/L		
	Total Organic Carbon (TOC)		6 - 100	mg/L		