

Proficiency Testing - PCBs in Water (1), priority pollutant

Art. ID ERA-832S
Unit each
Deliverydetails No Dangerous Good /not restricted

Description

One 2 mL flame-sealed ampule yields up to 2 liters after dilution. Contains a different Aroclor, randomly selected from the list below at 1-15 µg/L.

Text/Information	Analyte/Parameter	CAS number	Concentration/Value	Unit	Method	Source
	Aroclor 1016	[12674-11-2]	3,8 - 13	µg/L		
	Aroclor 1221	[11104-28-2]	1,0 - 15	µg/L		
	Aroclor 1232	[11141-16-5]	1,4 - 4	µg/L		
	Aroclor 1242	[53469-21-9]	3,8 - 13	µg/L		
	Aroclor 1248	[12672-29-6]	1,5 - 5,5	µg/L		
	Aroclor 1254	[11097-69-1]	1,7 - 5,5	µg/L		
	Aroclor 1260	[11096-82-5]	1,6 - 5	µg/L		