

Proficiency Testing - Haloacetic Acids (HAA), potable water

Art. ID	ERA-852
Unit	each
Deliverydetails	ADR Excepted Quantity (EQ) / AIR Excepted Quantity (EQ) UN2398 II E2

Description

One 2 mL flame-sealed ampule yields up to 2 liters after dilution. Includes all the analytes below at 5-50 µg/L.

Text/Information	Analyte/Parameter	CAS number	Concentration/Value	Unit	Method	Source
	Bromochloroacetic acid	[5589-96-8]	5,0 - 50	µg/L		
	Dibromoacetic acid	[631-64-1]	5,0 - 50	µg/L		
	Dichloroacetic acid	[79-43-6]	5,0 - 50	µg/L		
	Bromoacetic acid	[79-08-3]	5,0 - 50	µg/L		
	Chloroacetic acid	[79-11-8]	10,0 - 50	µg/L		
	Trichloroacetic acid	[76-03-9]	5,0 - 50	µg/L		