

Benzoic acid - DSC calibration standard

Art. ID LGC2606
Unit 500 mg
Deliverydetails No Dangerous Good /not restricted

Description

This certified reference material is intended for use in the calibration of differential scanning calorimeters and similar instruments. Certified values:

Text/Information	Analyte/Parameter	CAS number	Concentration/Value	Unit	Method	Source
	Enthalpy of Fusion		17,98	kJ/mol		
	Melting temperature		122,35	°C		