

## **Lead - DSC calibration standard**

Art. ID	LGC2608
Unit	500 mg
Deliverydetails	No Dangerous Good /not restricted

### Description

This certified reference material is intended for use in the calibration of differential scanning calorimeters and similar instruments. Certified values:

Text/Information	Analyte/Parameter	CAS number	Concentration/Value	Unit	Method	Source
	Enthalpy of Fusion		4,765	kJ/mol		
	Melting temperature		327,47	°C		