

Tetra-CDD Column Performance Check Solution

Art. ID	CIL-ED-935-A
Unit	0,5 mL
Deliverydetails	De Minimis UN1920 III E1

Description

Chemical purity: 98%

Text/Information	Analyte/Parameter	CAS number	Concentration/Value	Unit	Method	Source
	2,3,7,8-Tetrachlorodibe nzo-p-dioxin	[1746-01-6]	100	ng/mL		
	1,2,3,4-Tetrachlorodibe nzo-p-dioxin	[30746-58-8]	100	ng/mL		
	1,4,7,8-Tetrachlorodibe nzo-p-dioxin	[n/a]	100	ng/mL		
	1,2,3,7/1,2,3,8-Tetrach lorodibenzo-p-dioxin pa ir		100	ng/mL		
	1,2,7,8-Tetrachlorodibe nzo-p-dioxin	[34816-53-0]	200	ng/mL		
	2,3,7,8-Tetrachlorodibe nzo-p-dioxin (13C12,99%)	[1746-01-6]/unl abelled	250	ng/mL		
	37Cl4-2,3,7,8-Tetrachlo rodibenzo-p-dioxin	[1746-01-6]/unl abelled	7	ng/mL		
	1,2,3,4-Tetrachlorodibe nzo-p-dioxin (13C12,99%)	[114423-99-3]	500	ng/mL		