

## **instrument Check Standard 6 for Ion Chromatography in Water**

Art. ID LM24-SP-ICMIX6-100  
Unit 125 mL  
Deliverydetails No Dangerous Good /not restricted

Text/Information	Analyte/Parameter	CAS number	Concentration/Value	Unit	Method	Source
	Fluoride (F-)		20	mg/L		
	Nitrite (NO <sub>2</sub> -)N		25	mg/L		
	Chloride (Cl-)		50	mg/L		
	Bromide (Br-)		100	mg/L		
	HPO <sub>4</sub> -2		150	mg/L		
	Sulfate (SO <sub>4</sub> 2-)		150	mg/L		