

## QuiKResponse - PFAS Drinking Water

Art. ID ERA-735QR  
Unit each

### Description

One 2ml flame sealed ampule yields in excess of 1.5 L after dilution. Use with EPA method 537. The diluted standard will contain 6-8 analytes in each lot selected from the list below at 50-500 ng/L. Except PFBS will be in range of 100-1000ng/L.

Text/Information	Analyte/Parameter	CAS number	Concentration/Value	Unit	Method	Source
	N-ethyl perfluorooctane sulfonamidoacetic acid	[2991-50-6]	50 - 500	ng/L		
	N-methyl perfluorooctane sulfonamidoacetic acid	[2355-31-9]	50 - 500	ng/L		
	Perfluorobutanesulfonic acid (PFBS)	[375-73-5]	100 - 1000	ng/L		
	Perfluorodecanoic acid (PFDA)	[335-76-2]	50 - 500	ng/L		
	Perfluorododecanoic acid (PFDoA)	[307-55-1]	50 - 500	ng/L		
	Perfluoroheptanoic acid (PFHpA)	[375-85-9]	50 - 500	ng/L		
	Perfluorohexanesulfonic acid (PFHxS)	[355-46-4]	50 - 500	ng/L		
	Perfluorohexanoic acid (PFHxA)	[307-24-4]	50 - 500	ng/L		
	Perfluorononanoic acid (PFNA)	[375-95-1]	50 - 500	ng/L		
	Perfluorooctanesulfonic acid (PFOS)	[1763-23-1]	50 - 500	ng/L		
	Perfluorooctanoic acid (PFOA)	[335-67-1]	50 - 500	ng/L		
	Perfluorotetradecanoic acid (PFTA)	[376-06-7]	50 - 500	ng/L		
	Perfluorotridecanoic acid (PFTTrDA)	[72629-94-8]	50 - 500	ng/L		
	Perfluoroundecanoic acid (PFUnA)	[2058-94-8]	50 - 500	ng/L		