

## **Sulfur and Chlorine in Lube Oil CRM (S=0.3%, CL=0.03%)**

Art. ID	AR-3513
Unit	0,1 L
Deliverydetails	No Dangerous Good /not restricted

### Description

The intended use of this standard is for the calibration and verification of Sulfur and Chlorine by various valid test methods. This standard was produced gravimetrically using high purity materials, with balances calibrated and checked by precision NIST traceable weights, and verified by EDXRF.

Text/Information	Analyte/Parameter	CAS number	Concentration/Value	Unit	Method	Source
	Sulfur (S)	[7704-34-9]	0,03	%	ASTM D7751 Chlorine and Sulfur by EDXRF scan	
	Chlorine (Cl)	[7782-50-5]	0,03	%	ASTM D7751 Chlorine and Sulfur by EDXRF scan	