

0.25 mg/L C NIST Sucrose Calibration Standard (Anatel PAT700)

| | |
|-----------------|-----------------------------------|
| Art. ID | REA-ISTOC1002 |
| Unit | 60 mL |
| Deliverydetails | No Dangerous Good /not restricted |

Description

This product is prepared gravimetrically on a weight/volume basis to a specification of $\pm 2\%$. Reagecon holds ISO 17025 accreditation (A2LA Ref: 6739.02) for calibration of laboratory balances. The concentration of this standard is verified using a high performance calibrated Total Organic Carbon Analyser (TOC Instrument) using Reagecon's ISO 17025 accredited (A2LA Ref: 6739.03) test method. Current Certificates are 17025 approved / Traceable to a NIST standard

| Text/Information | Analyte/Parameter | CAS number | Concentration/Value | Unit | Method | Source |
|------------------|--------------------------------|------------|---------------------|------|--------|--------|
| | Total Organic Carbon (T OC) | | 0,25 | mg/L | | |