

0.5 mg/L C from USP Sucrose System Suitability Standard (Anatel PAT700)

Art. ID	REA-ISTOC1006
Unit	60 mL
Deliverydetails	No Dangerous Good /not restricted

Description

This product is prepared gravimetrically on a weight/volume basis to a specification of $\pm 2\%$. Reagecon holds ISO 17025 accreditation (A2LA Ref: 6739.02) for calibration of laboratory balances. The concentration of this standard is verified using a high performance calibrated Total Organic Carbon Analyser (TOC Instrument) using Reagecon's ISO 17025 accredited (A2LA Ref: 6739.03) test method.. Current Certificates are 17025 approved / Traceable to a NIST standard

Text/Information	Analyte/Parameter	CAS number	Concentration/Value	Unit	Method	Source
	Total Organic Carbon (T OC)		0,5	mg/L		